



CERTIFICATION by CUSTODIAN of RECORDS

Tucson Police Department Crime Laboratory  
270 S. Stone Avenue Tucson, AZ 85701-1917

I hereby certify that these documents are a true and correct copy of the periodic maintenance\* for the Intoxilyzer 8000 Serial Number 80-602265 maintained by the Tucson Police Department Crime Laboratory pursuant to the requirements of the Arizona Department of Public Safety for a quality assurance program.

TS4  
3.15.2011

A quality assurance specialist (QAS) tested the instrument on the following dates. The acceptable accuracy limits, as defined in A.A.C. R13-10-104.A.2, for a 0.100AC solution are 0.090AC to 0.110AC.

Date: 03/10/2011 Cal 1 0.101 AC Cal 2 0.100 AC

Between the dates of 02/10/2011 and 03/10/2011 the instrument was operating properly and accurately.

The instrument was taken out of service for repair on N/A.  
The instrument was returned to service on N/A.

Cheryl K. Straub


03/15/2011

ADPS QAS Permit # 33408

SUBSCRIBED AND SWORN to before me this 15, day of March, 20 11.

Notary Public

\*Includes documents specified in A.A.C. R13-10-104.A.5



Pursuant to section 28-1327, Arizona Revised Statutes, this document is a certified duplicate of the information contained in computer storage devices of the Tucson Police Department.

197  
CS

THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

TS4  
3.15.2011QA SPECIALIST Cheryl STRAUB AGENCY Tucson Police DepartmentDATE 03/10/2011 TIME 07:51INTOXILYZER SERIAL # 80-002265 LOCATION TEAM 5

(✓) 1. Display reads "PUSH BUTTON TO START"

## DIAGNOSTIC TESTS

(✓) 1. Clock time check.

(✓) 2. Date check.

## OPERATIONAL TESTS

(✓) 1. Alcohol-free subject test result 0. 000 AC.

(✓) 2. Error recognition logic system functioning.

Not a Successfully Completed Test Sequence printed

(✓) 3. Proper sample recognition system.

Not a Successfully Completed Test Sequence printed

Deficient sample printed.

(✓) 4. Standard Calibration Check standard 0. 100 ACResult: 0. 100 ACInstrument is operating properly and accurately. YES ✓ NO       

## COMMENTS

---

---

---

---

SIGNATURE

Cheryl Straub

287  
C8

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number : 80-002265

03/10/2011

07:53:17

BS4

Standard Lot#: 20910100A1

Last Changed By: C. STRAUB

#50772

QAS: C. STRAUB

#50772

TUCSON PD

Operator: C. STRAUB

#50772

TUCSON PD

Subject: SQAP, SQAP

DOB: 11/11/1111

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:54:03
Diagnostic Test	Pass	07:54:30
Air Blank	0.000	07:54:58
0.100 Cal Check	0.101	07:55:16
Air Blank	0.000	07:55:46
Subject Test	0.000	07:56:12
Air Blank	ABT*	07:56:19
Air Blank	0.000	07:56:47

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

Pursuant to section 14-127, Arizona Revised Statutes, this document, including the duplicate of the information contained in computer storage devices of the Tucson Police Department.

397  
CS

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number : 80-002265

03/10/2011

07:56:50

fy

Standard Lot#: 20910100A1

Last Changed By: C. STRAUB

#50772

QAS: C. STRAUB

#50772

TUCSON PD

Operator: C. STRAUB

#50772

TUCSON PD

Subject: SQAP, SQAP

DOB: 11/11/1111

Sex: F

Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	07:57:26
Air Blank	IPS**	07:57:33

\*RFI Detect

\*\*Improper Sample

487  
CS

INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number : 80-002265  
03/10/2011

07:57:39

PM

Standard Lot#: 20910100A1  
Last Changed By: C. STRAUB #50772

QAS: C. STRAUB #50772  
TUCSON PD

Operator: C. STRAUB #50772  
TUCSON PD

Subject: SQAP, SQAP  
DOB: 11/11/1111  
Sex: F Weight: 111

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:58:28
Diagnostic Test	Pass	07:58:55
Air Blank	0.000	07:59:23
0.100 Cal Check	0.101	07:59:42
Air Blank	0.000	08:00:11
Subject Test	0.179*	08:03:19
Air Blank	0.000	08:03:53
Air Blank	0.000	08:04:20
Subject Test	XXX**	08:04:35
Air Blank	0.000	08:05:09

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period



587  
CS


INTOXILYZER 8000  
Location: TUCSON POLICE DEPT  
Serial Number: 80-002265  
03/10/2011 08:05:55

104

-----  
Standard Lot#: 20910100A1  
Last Changed By: C. STRAUB #50772  
-----

QAS: C. STRAUB #50772  
TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	08:06:22
Cal Check	0.100	08:06:41
Air Blank	0.000	08:07:11

  
Operator's Signature

## EXHIBIT G-2

687  
08THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY  
A.A.C. R13-10-104(A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000154  
3.15.2011

STANDARD CALIBRATION CHECK PROCEDURE

QA Specialist Cheryl STRAUB AGENCY Tucson Police Department  
DATE 03/10/2011 TIME 07:51  
Intoxilyzer Serial # 80-002245 Location TEAM 5

- ( ✓ ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0. 100 AC.  
OR  
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- ( ✓ ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- ( ✓ ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- ( ✓ ) 4. Air blank completed.
- ( ✓ ) 5. Calibration check completed. Test results 0. 101 AC.
- ( ✓ ) 6. Air blank completed.
- ( ✓ ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Cheryl Str

DPS Form Exh G-2 (Rev. 05-01)

787  
C8

INTOXILYZER 8000

Location: TUCSON POLICE DEPT

Serial Number: 80-002265

03/10/2011

07:51:53

154


-----  
Standard Lot#: 20910100A1

Last Changed By: C. STRAUB #50772  
-----

QAS: C. STRAUB #50772

TUCSON PD  
-----

Test	g/210L	Time
Air Blank	0.000	07:52:20
Cal Check	0.101	07:52:39
Air Blank	0.000	07:53:09



Operator's Signature